



Sodium thiosulfate (stabilized) digital titrator cartridge, 2.26 N

Termékszám:	2686901
HUF Ár (nettó):	Elérhetőségünk
Szállítás 2 héten belül	

For determination of Hypochlorite (Bleach) by Digital Titrator titration. Hach Method 10100. Range: 5 - 15% Cl₂ (50 - 150 g/L Cl₂). Stabilized sodium thiosulfate formulation. 2.26 N Digits x 0.5 = g/L Cl₂ in 0.2 mL sample

Accurate results - when and where you need them

Digital Titrator methods use a high-precision dispensation device and cartridges of concentrated titrant to allow analyses in the lab, the plant or the field.

Fast and Flexible

Interchangeable cartridges allow you to change titrants in seconds. No cleaning or rinsing burets.

Versatile

More than 50 parameter/range combinations are available

Economical

Concentrated titrants allow 40-100 typical titrations per cartridge.

Műszaki adatok

Méréstartomány:	50 - 150 g/L Cl ₂
Módszer név:	Iodometric Method
Paraméter:	Hypochlorite
Platform:	Digitális titrátor
Tesztek száma:	40 - 100
Titrló:	2.26 N Na ₂ S ₂ O ₃