



Chlorine (Free and Total) Reagent Set, DPD FEAS, Digital Titrator

Termékszám:

2445300

HUF Ár (nettó):

Elérhetőségünk

Szállítás 2 héten belül

Veszélyes



For determination of Chlorine (Free and Total) by Ferrous Ethylenediammonium Sulfate Digital Titration, using DPD indicator. Range: Up to 3.00 mg/L Cl_2 .

Reagent Set includes: DPD Free Chlorine Reagent Powder Pillows, DPD Total Chlorine Reagent Powder Pillows, Ferrous Ethylenediammonium Sulfate Titration Cartridge, 0.00564 N. Approximately 40-100 tests.

Boric Acid free!

Accurate results - when and where you need them

Digital Titrator methods use a high-precision dispensation device and cartridges of concentrated titrant to allow analyses in the lab, the plant or the field.

Fast and Flexible

Interchangeable cartridges allow you to change titrants in seconds. No cleaning or rinsing burets.

Versatile

More than 50 parameter/range combinations are available

Economical

Concentrated titrants allow 40-100 typical titrations per cartridge.

Műszaki adatok

Méréstartomány: 0 - 3.00 mg/L Cl_2

Módszer név: DPD-FEAS

Tesztek száma: 40 - 100